

Office: 802-540-0148 | Fax: 802-540-0147 480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis

Company: Old Growth Vermont Sample ID: Blue Tsi Fly

Lot: CLTV0058-001

V0058-001

Date Analyzed: 1/27/2023

Report Date: 1/30/2023

Date Sampled: N/A

Analyst: 050

Grower License #: CLTV0058

Customer ID: 221024-2

Date Received: 1/20/2023

Matrix: Flower

Report ID: C230120AM

Cannabinoid Summary

			Calmadinolu S
Cannabinoid Profile	LOQ (mg/g)	Concentration (mg/g)	Weight (%)
CBDVA	0.0005	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
CBDV	0.0012	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
CBDA	0.0008	0.70	0.07
CBGA	0.0008	2.11	0.21
CBG	0.0019	0.80	0.08
CBD	0.0019	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
THCV	0.0021	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
CBN	0.0013	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Δ9-ТНС	0.0020	4.16	0.42
Δ8-THC	0.0019	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
THC-A	0.0034	173.08	17.31
СВС	0.0024	0.47	0.05
Total THC		155.95	15.60
Total CBD		0.62	0.06
Total Cannabir	noids	181.33	18.13

15.6% 0.06%

Total THC Total CBD

18.13% 0.42%

Total
Cannabinoids Δ9-THC

12.35%
Percent
Moisture

1:0
THC:CBD
Ratio

Cannabinoids Methodology: High Performance Liquid Chromatography (HPLC) using PerkinElmer FLEXAR™ with Photo Diode Array Detector (PDA)

LOQ = The lowest quantity that this method can reliably detect. Any cannabinoid that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

Measurement of Uncertainty (MU): the parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the particular quantity subject to measurement. $\Delta 9\text{-THC MU} = \pm 0.005\%$ Total THC MU = $\pm 0.007\%$

All other cannabinoid MU values are available upon request.

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the Certified by: samples as received.



Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

(802) 540-0148 laboratory@biadiagnostics.com Certificate Registration Number: CL_50_2021_002



Customer ID: 221024-2

Grower License #: CLTV0058

Office: 802-540-0148 | Fax: 802-540-0147 480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis

Company: Old Growth Vermont Sample ID: Harvest Lot Composite

Lot: CLTV0058-001 Report Date: 1/26/2023 Matrix: Flower Date Analyzed: 1/26/2023

Date Sampled: N/A Analyst: 018

Date Received: 1/20/2023 Report ID: C230120AS-2 Amendment to C230120AS

Pathogen Summary

Target Pathogens	Method	LOD (cfu/g)	Result (cfu/g)
Aspergillus - flavus, fumigatus, niger, terreus	Aspergillus AOAC PTM No. 032104	5	<lod< td=""></lod<>
STEC	STEC Virx AOAC PTM No. 121203	5	<lod< td=""></lod<>
Salmonella spp.	Salmonella II AOAC PTM No. 010803	5	<lod< td=""></lod<>



Test Methodology: Bio-Rad IQ-Check PCR Kits

cfu/g = colony forming units per gram

LOD = The lowest quantity that this method can reliably detect. Any microbial growth that was not detected is assumed to be less than the stated LOD (<LOD).

Reagent Blanks: <LOD for all analytes

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the samples as received.

Certified by:

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

(802) 540-0148 laboratory@biadiagnostics.com



Office: 802-540-0148 | Fax: 802-540-0147 480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis

Company: Old Growth Vermont Sample ID: Harvest Lot Composite

Lot: CLTV0058-001 Report Date: 1/26/2023

Matrix: Flower Date Analyzed: 1/25/2023

Customer ID: 221024-2 Date Sampled: N/A Analyst: 45

Grower License #: CLTV0058 Date Received: 1/20/2023 Report ID: C230120AS-2 Amendment to C230120AS

Pesticides/Mycotoxins Summary

Category II Residual Pesticide	LOQ (ppm)	Concentration (ppm)	
Abamectin	0.0100	<loq< td=""></loq<>	
Acephate	0.0010	<loq< td=""></loq<>	
Acequinocyl	0.0010	<loq< td=""></loq<>	
Azoxystrobin	0.0010	<loq< td=""></loq<>	
Bifenazate	0.0010	<loq< td=""></loq<>	
Bifenthrin	0.0010	<loq< td=""></loq<>	
Carbaryl	0.0010	<loq< td=""></loq<>	
Cypermethrin	0.0100	<loq< td=""></loq<>	
Etoxazole	0.0010	<loq< td=""></loq<>	
Imidacloprid	0.0010	<loq< td=""></loq<>	
Myclobutanil	0.0010	<loq< td=""></loq<>	
Pyrethrin I	0.0010	<loq< td=""></loq<>	
Pyrethrin II	0.0010	<loq< td=""></loq<>	
Spinosyn A	0.0010	<loq< td=""></loq<>	
Spinosyn D	0.0010	<loq< td=""></loq<>	

Category II Mycotoxin	LOQ (ppm)	Concentration (ppm)
Ochratoxin A	0.0020	NOT TESTED
Aflatoxin B1	0.0002	NOT TESTED
Alfatoxin B2	0.0010	NOT TESTED
Alfatoxin G1	0.0002	NOT TESTED
Alfatoxin G2	0.0010	NOT TESTED

Category I Residual Pesticide	LOQ (ppm)	Concentration (ppm)
Chlorpyrifos	0.0010	<loq< th=""></loq<>
Imazalil	0.0010	<loq< th=""></loq<>



9.27%

Percent Moisture

LOQ = The lowest quantity this method can reliably detect. Any pesticide or mycotoxins that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

ppb = parts per billion

Pesticides/Mycotoxin Methodology: Liquid Chromatography with Tandem Mass Spectrometry using PerkinElme QSight® LX50 UHPLC and QSight 220 Mass Spectrometer

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

Certified by: Luke E.M

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context.

Results apply to the samples as received.

(802) 540-0148 laboratory@biadiagnostics.com



Customer ID: 221024-2

Grower License #: CLTV0058

Office: 802-540-0148 | Fax: 802-540-0147 480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis

Company: Old Growth Vermont Sample ID: Blue To Fly

Lot: N/A

Matrix: Flower Date Sampled: N/A

Date Received: 1/20/2023

Report Date: 1/27/2023 **Date Analyzed:** 1/25/2023

Analyst: 035

Report ID: C230120AM

Terpenes Summary

Terpene	LOQ (mg/g)	Results (mg/g)	Weight (%)
α- Pinene	0.010	1.083	0.108
Camphene	0.010	0.197	0.020
β-Myrcene	0.010	4.256	0.426
b-Pinene	0.010	1.663	0.166
3-Carene	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
α-Terpinene	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Limonene	0.010	6.265	0.627
ρ-Cymene	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Ocimene	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Eucalyptol	0.010	0.154	0.015
Y-Terpinene	0.010	0.028	0.003
Terpinolene	0.010	0.140	0.014
Linalool	0.010	0.178	0.018
Isopulegol	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Geraniol	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Caryophyllene	0.010	4.072	0.407
α-Humulene	0.010	1.749	0.175
Trans-Nerolidol	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Cis-Nerolidol	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Guaiol	0.010	<loq< th=""><th><loq< th=""></loq<></th></loq<>	<loq< th=""></loq<>
Caryophyllene Oxide	0.010	0.020	0.002
α-Bisabolol	0.010	0.029	0.003
Total Terpenes		19.834	1.984

12.35%

LOQ = The lowest quantity this method can reliably detect. Any terpene that was not detected is assumed to be less than the stated LOQ (<LOQ).

Percent Moisture Terpene Methodology: Headspace Sampler, Gas Chromatography-Mass Spectrometry (GC-MS), using Perkin Elmer Clarus® SQ8 GC MS

Reagent Blanks: < LOQs for all analytes

All results reflect dry weight of material, based on % moisture of the sample.

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.



This report shall not be reproduced except in full without approval of the laboratory.

This is to provide assurance that parts of a report are not taken out of context. Results apply to the samples as received.

Certified by:

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

Luke E.M

(802) 540-0148 laboratory@biadiagnostics.com